

# Certificate of Analysis

Apex Biological Indicator (Reorder # NAS-152)  
for Gaseous Hydrogen Peroxide

Lot #: **N0975**

Manufacture: 2015 APRIL 02      Expiration: 2016 JANUARY 31

Indicator: *Geobacillus stearothermophilus* 7953 <sup>(1)</sup>

Mean population: **2.3** x 10<sup>6</sup> CFU per stainless steel carrier <sup>(2)</sup>

Storage conditions: 2-8°C; less than 50% RH; move to ambient conditions ≥ 1 hr before use.

Shipping conditions: Ambient temperatures; cold pack and desiccant may be used to moderate conditions during shipping.

## Resistance Characteristics:

D-value <sup>(3)</sup>: **0.7** minutes in 2mg/L gaseous H<sub>2</sub>O<sub>2</sub>

Units are manufactured in compliance with Mesa Laboratory, Bozeman Manufacturing Facility's quality standards and ISO 11138-1 guidelines and all appropriate subsections.

Purity: No evidence of contaminants using standard plate count techniques.

Incubate at 55-60°C for 7 days. The recommended growth medium is Soybean Casein Digest Medium (SCDM) / Tryptic Soy Broth (TSB).

This product is for Industrial Use Only.

Disposal: Treat as non-pathogenic material and sterilize (steam, EtO, etc) or incinerate before discarding.

<sup>(1)</sup> Culture is traceable to a recognized culture collection identified in USP and ISO 11138.

<sup>(2)</sup> Heat shock population determined at 95-100°C for 15 minutes

<sup>(3)</sup> Resistance was calculated using Fraction Negative method. The D-value is reproducible only when exposed and cultured under the exact conditions used to obtain results reported here.

Certified By: \_\_\_\_\_

Quality Representative



Bozeman Manufacturing Facility  
10 Evergreen Drive  
Bozeman, MT 59715  
T: 303/987-8000 • F: 406/585-9219  
www.mesalabs.com

Rev.1  
423MAR2015  
Part #: 36.7008-10

## BIOLOGICAL INDICATOR INSTRUCTIONS -- NAS-152

### PROCESS EVALUATION

•Place biological indicators in locations previously determined to be the most difficult to sterilize. Areas experiencing minimal gas flow or poor gas distribution include enclosure corners, areas in and around equipment, and locations among disposable materials to be used in the enclosure. Note that the inoculated side of the carrier faces the printed label on the Tyvek pouch, therefore **the printed side should face outward during a process cycle.**

•Validation and mapping processes generally require multiple indicators at numerous sites in an enclosure.

•Conduct the sterilization and aeration cycle.

•Remove the indicators and deliver them, plus one or more unexposed control indicators, to the laboratory for sterility testing. Culturing of exposed indicators should be conducted as soon as possible following removal from the enclosure being tested.

### CULTURING PROCEDURES

•Culture in a laminar flow hood. Observe strict aseptic technique at all times. Minimally, sterile gloves should be worn. Include donning hoods, masks, and gowns as appropriate for the facility and circumstances.

•Aseptically open the Tyvek pouch by cutting with sterile scissors or peeling apart at the end with the package offset.

•Using sterile forceps, withdraw the carrier and place in a tube containing sterile Soybean Casein Digest Medium (SCDM) / Tryptic Soy Broth (TSB).

•Aseptically culture the control carrier(s) last.

•Select one or more tubes of the same lot of culture medium to serve as negative controls.

•Incubate test and control tubes for 7 days at 55-60°C. Observe daily for evidence of growth (turbidity).

### INTERPRETATION

•Turbidity:

For test indicators, turbidity suggests that the sterilization was incomplete and that at least one spore survived the process.

For positive control indicators, turbidity indicates that viable spores were present and capable of outgrowth in the culture medium used.

In negative control tubes, turbidity indicates that viable organisms may have been present in the growth medium. Contact your supplier.

•No turbidity:

For test indicators, lack of turbidity indicates sterilization was complete and no spores survived the process.

In negative control tubes, lack of turbidity indicates no other viable organisms were present in the culture medium.

For positive control indicator, no turbidity suggests no viable organisms were present on the carrier or that the media may be inhibiting the outgrowth of the test organism. Contact your supplier.

**LIMITATION OF LIABILITY AND INDEMNITY:** In no event, whether as a result of breach of contract, warranty or tort (including negligence and strict liability) shall Mesa Labs or its suppliers be liable for any consequential or incidental damages including, but not limited to loss of profits or revenues, loss of use of the Products or any associated equipment, loss of the Buyer's Products, damage to associated equipment, cost of capital, cost of substitute products, facilities, service or replacement power, downtime cost, caused by such Products, or claims of the users for such damages. Buyer for itself, its successors and assigns, hereby agrees to indemnify Mesa Labs and to hold Mesa Labs harmless from any and all liability for such consequential or incidental damages. The responsibility of Mesa Labs for damages due to injuries to employees of the Buyer or ultimate user of the Product, caused by the Product, shall be limited to repair or replacement of the item, at the option of Mesa Labs. The Buyer agrees to indemnify Mesa Labs and hold Mesa Labs harmless from any further damages, indemnity or contribution. Mesa Labs liability for any claim of any kind, including performance or breach thereof, or from the Products to Services furnished hereunder, shall in no case exceed the price of the specified Product, system, component or service which gives rise to the claim.