


# ProChem OK - Steam Sterilization Multi-Spot Indicator



Catalog # CI-OK 250 Indicators / bag 3 year expiry

For International Use Only

Stated Values (As determined in a steam sterilization resistometer)		
Spot	121° C	134° C
1	8 minutes	3.25 minutes
2	11 minutes	4.5 minutes
3	17 minutes	7 minutes
4	30 minutes	11 minutes

- Pre-vacuum sterilizers operating at 121°C and 134°C
- Multiple spots with varied timing
- Single indicator suitable for multiple cycle lengths