MesaLabs

CERTIFICATE OF CONFORMANCE

Mesa Laboratories, Inc.
625 Zoot Way, Bozeman, MT 59718
Phone: (303) 987-8000 www.mesalabs.com

Product Name:	MESA ProChem SSW Integrator
Catalog No.:	77-05-CI-SSW
Lot Number:	217
Expiration Date:	2022-02-01
Product Description:	Mesa ProChem SSW is a single-use Type 5 integrating indicator.
Indications for Use:	The integrating indicator is designed to chemically react over time with the critical parameters of steam sterilization cycle within a specified tolerance. The integrating indicator strip is intended to be placed in each pack, pouch, container, tray or other containment device to function as an independent monitor of critical parameters for the following sterilization cycles:

Steam Sterilization Cycles:
- 250°F/121°C, 30 minutes Gravity
- 270°F/132°C, 4 minutes Dynamic Air Removal
- 270°F /132°C, 15 minutes Gravity
- 275°F/135°C, 3 minutes Dynamic Air Removal
- 275°F/135°C, 10 minutes Gravity

Steam Sterilization Cycles (IUSS):
- 270°F/132°C, 4 minutes Dynamic Air Removal
- 270°F/132°C, 3 minutes Gravity
- 270°F/132°C, 10 minutes Gravity
- 275°F/135°C, 3 minutes Dynamic Air Removal
- 275°F/135°C, 3 minutes Gravity
- 275°F/135°C, 10 minutes Gravity

The manufacturer of this product certifies it conforms to all specifications whether physical, dimensional or performance related and that test reports or process certifications are on file.

This certificate is issued to confirm that the product listed above has been tested by Mesa Labs and satisfies performance requirements after exposure to saturated steam when tested under the conditions described in ANSI/AAMI/ISO 11140-1 for Type 5 process indicators.

Certified by: _____ 20Dec2017
Quality Representative Date