

PRODUCT CHANGE NOTIFICATION

11 April 2019

Notification ID: CN-190401

Notification Type: Product Labeling Change

Dear Customer,

We would like to take this opportunity to notify you of labeling changes that will be made to one of our Spore Strip catalog numbers. This change applies to the following Mesa catalog number:

- 5-1000C Spore Strip Combined - Control Log6/5 1000-ct bag

The change is being made in order to align our primary labeling with the requirements of ISO 11138-1:2017 Section 4.3.1. The primary packaging and labeling of this catalog number will be changed as follows:


- We will no longer use glassine with any pre-printed labeling information. All primary labeling contents will be printed by Mesa during packaging of the spore strips.
- The title "Bacterial Spore Test Strip" will be changed to "Bacterial Spore Control Strip".
- Fonts will be changed.
- The eye mark and the word "PEEL" will now be printed in black instead of red.
- Lot number and expiration will now be printed on the front of the glassine instead of the back and be printed in red ink.

A sample label is provided below. Mesa anticipates that this change will take effect in June 2019 but it is dependent on the volume of customer orders for this item.

Spore Strip Combined - Control primary labeling



As always, if you have any questions, please contact your Mesa Laboratories Representative.



 Robert Bradley, Director — Manufacturing

11 April 2019

 Date



 Nicole Dewald, Manager — Quality Assurance

11 APR 2019

 Date